



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2019/04/03

REF TEK 147A-01/1/V

S/N:82010

P/N: 106293-02

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Vertical
1.00	0.9984
10.00	1.0000
20.00	1.0041
30.00	1.0086
40.00	1.0121
50.00	1.0136
60.00	1.0132
70.00	1.0114
80.00	1.0087
90.00	1.0060
100.00	1.0038
120.00	1.0037
140.00	1.0136
160.00	1.0330

OUTPUT SENSITIVITY

Vertical	
Output Sensitivity (V/g)	2.501

* g = 9.80 m/s²