



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2019/04/03

REF TEK 147A-01/1/V

S/N:82011

P/N: 106293-02

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Vertical
1.00	1.0002
10.00	1.0000
20.00	0.9995
30.00	0.9987
40.00	0.9976
50.00	0.9965
60.00	0.9954
70.00	0.9945
80.00	0.9942
90.00	0.9947
100.00	0.9964
120.00	1.0053
140.00	1.0245
160.00	1.0501

OUTPUT SENSITIVITY

Vertical	
Output Sensitivity (V/g)	2.500

* g = 9.80 m/s²