



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2019/09/27

REF TEK 147A-01/1/V

S/N:82014

P/N: 106293-02

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Vertical
1.00	0.9996
10.00	1.0000
20.00	1.0011
30.00	1.0025
40.00	1.0036
50.00	1.0039
60.00	1.0033
70.00	1.0020
80.00	1.0002
90.00	0.9987
100.00	0.9983
120.00	1.0035
140.00	1.0201
160.00	1.0420

OUTPUT SENSITIVITY

Vertical	
Output Sensitivity (V/g)	2.502

* g = 9.80 m/s²