



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2019/09/27

REF TEK 147A-01/1/V
S/N:82018

P/N: 106293-02

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Vertical
1.00	1.0003
10.00	1.0000
20.00	0.9991
30.00	0.9977
40.00	0.9959
50.00	0.9942
60.00	0.9925
70.00	0.9914
80.00	0.9912
90.00	0.9921
100.00	0.9945
120.00	1.0052
140.00	1.0251
160.00	1.0474

OUTPUT SENSITIVITY

Vertical	
Output Sensitivity (V/g)	2.501

* g = 9.80 m/s²