



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2019/09/27

REF TEK 147A-01/1/V

S/N:82023

P/N: 106293-02

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Vertical
1.00	0.9996
10.00	1.0000
20.00	1.0009
30.00	1.0020
40.00	1.0028
50.00	1.0028
60.00	1.0019
70.00	1.0003
80.00	0.9981
90.00	0.9962
100.00	0.9953
120.00	0.9991
140.00	1.0144
160.00	1.0367

OUTPUT SENSITIVITY

Vertical	
Output Sensitivity (V/g)	2.499

* g = 9.80 m/s²