



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2019/09/27

REF TEK 147A-01/1/V

S/N:82025

P/N: 106293-02

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Vertical
1.00	1.0001
10.00	1.0000
20.00	0.9997
30.00	0.9992
40.00	0.9986
50.00	0.9980
60.00	0.9975
70.00	0.9974
80.00	0.9979
90.00	0.9993
100.00	1.0019
120.00	1.0123
140.00	1.0317
160.00	1.0531

OUTPUT SENSITIVITY

Vertical	
Output Sensitivity (V/g)	2.500

* g = 9.80 m/s²