



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2019/09/27

REF TEK 147A-01/1/V
S/N:82028

P/N: 106293-02

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Vertical
1.00	0.9999
10.00	1.0000
20.00	1.0002
30.00	1.0004
40.00	1.0007
50.00	1.0009
60.00	1.0009
70.00	1.0010
80.00	1.0012
90.00	1.0019
100.00	1.0032
120.00	1.0100
140.00	1.0249
160.00	1.0429

OUTPUT SENSITIVITY

Vertical	
Output Sensitivity (V/g)	2.501

* g = 9.80 m/s²