



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2019/09/27

REF TEK 147A-01/1/V

S/N:82030

P/N: 106293-02

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Vertical
1.00	0.9988
10.00	1.0000
20.00	1.0028
30.00	1.0056
40.00	1.0070
50.00	1.0071
60.00	1.0059
70.00	1.0042
80.00	1.0025
90.00	1.0015
100.00	1.0017
120.00	1.0076
140.00	1.0232
160.00	1.0427

OUTPUT SENSITIVITY

Vertical	
Output Sensitivity (V/g)	2.499

* g = 9.80 m/s²