



# REF TEK

SYSTEMS INC.

**The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.**

**Reftek Systems Inc. purchased the REF TEK product line in May 2020.**

**Reftek Systems will honour existing product warranties and extended warranties.**

**ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:**

**Sales & Administrative:** [sales@reftek.com](mailto:sales@reftek.com)

**Support & Engineering:** [support@reftek.com](mailto:support@reftek.com)

**Accounting:** [accounting@reftek.com](mailto:accounting@reftek.com)

**Address:** 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

**Phone:** +1-902-444-7650

**Fax:** +1-902-444-7651

DATE: 2019/09/27

REF TEK 147A-01/1/V  
S/N:82033

P/N: 106293-02

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

| Freq. (Hz) | Vertical |
|------------|----------|
| 1.00       | 0.9991   |
| 10.00      | 1.0000   |
| 20.00      | 1.0022   |
| 30.00      | 1.0045   |
| 40.00      | 1.0059   |
| 50.00      | 1.0060   |
| 60.00      | 1.0051   |
| 70.00      | 1.0036   |
| 80.00      | 1.0022   |
| 90.00      | 1.0016   |
| 100.00     | 1.0022   |
| 120.00     | 1.0097   |
| 140.00     | 1.0274   |
| 160.00     | 1.0476   |

OUTPUT SENSITIVITY

| Vertical                 |       |
|--------------------------|-------|
| Output Sensitivity (V/g) | 2.499 |

\* g = 9.80 m/s<sup>2</sup>