



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2019/09/27

REF TEK 147A-01/1/V

S/N:82034

P/N: 106293-02

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Vertical
1.00	0.9996
10.00	1.0000
20.00	1.0010
30.00	1.0023
40.00	1.0033
50.00	1.0038
60.00	1.0034
70.00	1.0023
80.00	1.0007
90.00	0.9996
100.00	0.9994
120.00	1.0053
140.00	1.0235
160.00	1.0481

OUTPUT SENSITIVITY

Vertical	
Output Sensitivity (V/g)	2.499

* g = 9.80 m/s²