



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2019/09/29

REF TEK 147A-01/1/V

S/N:82035

P/N: 106293-02

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Vertical
1.00	0.9992
10.00	1.0000
20.00	1.0016
30.00	1.0026
40.00	1.0024
50.00	1.0013
60.00	0.9998
70.00	0.9983
80.00	0.9973
90.00	0.9973
100.00	0.9988
120.00	1.0082
140.00	1.0285
160.00	1.0543

OUTPUT SENSITIVITY

Vertical	
Output Sensitivity (V/g)	2.499

* g = 9.80 m/s²