



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

DATE: 2019/09/27

REF TEK 147A-01/1/V

S/N:82036

P/N: 106293-02

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Vertical
1.00	0.9998
10.00	1.0000
20.00	1.0004
30.00	1.0010
40.00	1.0016
50.00	1.0019
60.00	1.0020
70.00	1.0019
80.00	1.0019
90.00	1.0023
100.00	1.0037
120.00	1.0117
140.00	1.0300
160.00	1.0524

OUTPUT SENSITIVITY

Vertical	
Output Sensitivity (V/g)	2.501

* g = 9.80 m/s²