



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651



Refraction Technology, Inc.
1600 Tenth Street, Suite A
Plano, Texas 75074

<http://www.reftek.com> [Tel:214-440-1265](tel:214-440-1265) Fax: 972-578-0045

Broadband Feedback Seismometer Parameters Table 151-120 Observer

Seismometer serial number	Mechanical parts' serial numbers	Period (sec)	Damp coefficient	Output voltage sensitivity (v.s/ m)	Calibration sensitivity (m/ s ² / A)
G08331VS	ud_G08331	120.0	0.707	1991.84	10.907
	ew_G08331	120.0	0.707	1990.62	10.6701
	ns_G08331	120.0	0.707	2000.21	10.5584

The values above are practically measured