The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Refraction Technology, Inc.
1600 Tenth Street, Suite A
Plano, Texas 75074
http://www.reftek.com Tel: 214-440-1265 Fax: 972-578-0045

151-30 Observer
Parameters Table

Seismometer Serial Number	Component	Period (sec)	Damping Coefficient	Output Voltage Sensitivity (V/m/s)	Calibration Sensitivity ($m/s^2/A$)
G09303VS	**Vertical**	30.19	0.715	2016.56	10.5525
	North/South	29.98	0.706	2009.32	10.8924
	East/West	30.09	0.714	2008.26	11.0001

Zeros: 2 radians

z1 = 0
z2 = 0

Poles: 4 radians

p1 = -0.14810 +0.14810i
p2 = -0.14810 -0.14810i
p3 = -177.72 +177.72i
p4 = -177.72 - 177.72i