



# REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

**Reftek Systems Inc. purchased the REF TEK product line in May 2020.**

**Reftek Systems will honour existing product warranties and extended warranties.**

**ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:**

**Sales & Administrative:** [sales@reftek.com](mailto:sales@reftek.com)

**Support & Engineering:** [support@reftek.com](mailto:support@reftek.com)

**Accounting:** [accounting@reftek.com](mailto:accounting@reftek.com)

**Address:** 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

**Phone:** +1-902-444-7650

**Fax:** +1-902-444-7651



Refraction Technology, Inc.  
 1600 Tenth Street, Suite A  
 Plano, Texas 75074

<http://www.reftek.com> Tel: 214-440-1265 Fax: 972-578-0045

## 151-30A Observer Parameters Table

Seismometer Serial Number	Component	Period (sec)	Damping Coefficient	Output Voltage Sensitivity (V/m/s)	Calibration Sensitivity (m/s <sup>2</sup> /A)
<b>G11308VS</b>	<b>Vertical</b>	30.2	0.704	1997.54	10.1194
	<b>North/South</b>	30.2	0.709	2005.46	9.9281
	<b>East/West</b>	30.2	0.708	1997.35	9.9701

Zeros: 3 radians

z1 = 0

z2 = 0

z3 = -1.8180e +001

Poles: 6 radians

p1 = 1.4807e -001 +1.4812e -001 i

p2 = -1.4807e -001 -1.4812e -002 i

p3 = -1.8033e +002 +1.8992e +002 i

p4 = -1.8033e +002 -1.8992e +002 i

P5 = -2.0958e +001

P6 = -7.72203e +002