The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Refraction Technology, Inc.
1600 Tenth Street, Suite A
Plano, Texas 75074
http://www.reftek.com Tel: 214-440-1265 Fax: 972-578-0045

151-30A Observer
Parameters Table

Seismometer Serial Number	Component	Period (sec)	Damping Coefficient	Output Voltage Sensitivity (V/m/s)	Calibration Sensitivity ($m/s^2/A$)
G11328VS	**Vertical**	29.9	0.702	1992.94	10.4495
	North/South	30.1	0.705	1987.45	10.3927
	East/West	29.9	0.703	1992.41	10.3945

Zeros: 3 radians
z1 = 0
z2 = 0
z3 = -1.8180e +001

Poles: 6 radians
p1 = -1.4807e -001 +1.4812e -001 i
p2 = -1.4807e -001 -1.4812e -002 i
p3 = -1.8033e +002 +1.8992e +002 i
p4 = -1.8033e +002 -1.8992e +002 i
P5 = -2.0958e +001
P6 = -7.72203e +002