



# REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

**Reftek Systems Inc. purchased the REF TEK product line in May 2020.**

**Reftek Systems will honour existing product warranties and extended warranties.**

**ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:**

**Sales & Administrative:** sales@reftek.com

**Support & Engineering:** support@reftek.com

**Accounting:** accounting@reftek.com

**Address:** 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

**Phone:** +1-902-444-7650

**Fax:** +1-902-444-7651

Trimble Navigation Limited  
 1600 Tenth Street  
 Suite A  
 Plano, Texas 75074  
 Tel: 214-440-1265  
 Fax: 972-578-0045  
[www.reftek.com](http://www.reftek.com)



## 151-60A Observer Parameters Table

Seismometer Serial Number	Component	Period (sec)	Damping Coefficient	Output Voltage Sensitivity (V/m/s)	Calibration Sensitivity (m/s <sup>2</sup> /A)
G10634VS	Vertical	61.0	0.705	1991.79	10.1090
	North/South	60.8	0.702	2015.10	10.0002
	East/West	60.9	0.697	1997.35	10.0719

**Zeros: 3 radians**

z1 = 0  
 z2 = 0  
 z3 = -1.817998e +001

**Poles: 6 radians**

p1 = -7.403687e -002 +7.405923e -002 i  
 p2 = -7.403687e -002 -7.405923e -002 i  
 p3 = -1.803307e +002 +1.899224e +002 i  
 p4 = -1.803307e +002 -1.899224e +002 i  
 p5 = -2.095807e +001  
 p6 = -7.720300e +002