



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651



REF TEK 151B-120 Observer Parameters Table

Seismometer Serial Number	Component	Period (sec)	Damping Coefficient	Output Voltage Sensitivity (V/m/s)	Calibration Sensitivity (m/s ² /A)
G14151VS	Vertical	119.84	0.701	1982.66	11.61
	North/South	119.82	0.705	2003.82	10.50
	East/West	119.87	0.701	1988.77	10.50

Zeros: 2 radians

$z1 = 0$

$z2 = 0$

Poles: 4 radians

Vertical	$p1 = -0.036753 + 0.037391 i$	$p3 = -222.144 + 222.144 i$
	$p2 = -0.036753 - 0.037391 i$	$p4 = -222.144 - 222.144 i$
North/South	$p1 = -0.036969 + 0.037190 i$	$p3 = -222.144 + 222.144 i$
	$p2 = -0.036969 - 0.037190 i$	$p4 = -222.144 - 222.144 i$
East/West	$p1 = -0.036744 + 0.037382 i$	$p3 = -222.144 + 222.144 i$
	$p2 = -0.036744 - 0.037382 i$	$p4 = -222.144 - 222.144 i$