



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651



REF TEK 151B-120 Observer Parameters Table

Seismometer Serial Number	Component	Period (sec)	Damping Coefficient	Output Voltage Sensitivity (V/m/s)	Calibration Sensitivity (m/s ² /A)
G14179VS	Vertical	119.99	0.699	2003.16	11.67
	North/South	120.70	0.703	1999.60	10.71
	East/West	119.96	0.699	2013.49	10.58

Zeros: 2 radians

$z1 = 0$

$z2 = 0$

Poles: 4 radians

Vertical	p1 = -0.036603 +0.037447 i	p3 = -222.144 +222.144 i
	p2 = -0.036603 -0.037447 i	p4 = -222.144 -222.144 i
North/South	p1 = -0.036596 +0.037022 i	p3 = -222.144 +222.144 i
	p2 = -0.036596 -0.037022 i	p4 = -222.144 -222.144 i
East/West	p1 = -0.036612 +0.037456 i	p3 = -222.144 +222.144 i
	p2 = -0.036612 -0.037456 i	p4 = -222.144 -222.144 i