



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651



REF TEK 151B-60 Observer Parameters Table

Seismometer Serial Number	Component	Period (sec)	Damping Coefficient	Output Voltage Sensitivity (V/m/s)	Calibration Sensitivity (m/s ² /A)
G12631VS	Vertical	60.77	0.7150	1994.92	11.6167
	North/South	60.34	0.7100	1985.37	10.7311
	East/West	60.49	0.7120	1983.78	10.8276

Zeros: 2 radians

z1 = 0

z2 = 0

Poles: 4 radians

Vertical	p1 = -0.073926 +0.072284 i	p3 = -222.144 +222.144 i
	p2 = -0.073926 -0.072284 i	p4 = -222.144 -222.144 i
North/South	p1 = -0.073932 +0.073328 i	p3 = -222.144 +222.144 i
	p2 = -0.073932 -0.073328 i	p4 = -222.144 -222.144 i
East/West	p1 = -0.073956 +0.072936 i	p3 = -222.144 +222.144 i
	p2 = -0.073956 -0.072936 i	p4 = -222.144 -222.144 i