



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651



REF TEK 151B-60 Observer Parameters Table

Seismometer Serial Number	Component	Period (sec)	Damping Coefficient	Output Voltage Sensitivity (V/m/s)	Calibration Sensitivity (m/s ² /A)
G12639VS	Vertical	60.57	0.707	2005.13	11.69
	North/South	60.41	0.707	1997.07	10.68
	East/West	60.68	0.708	2011.53	10.89

Zeros: 2 radians

z1 = 0
 z2 = 0

Poles: 4 radians

Vertical	p1 = -0.073340 +0.073362 i	p3 = -222.144 +222.144 i
	p2 = -0.073340 -0.073362 i	p4 = -222.144 -222.144 i
North/South	p1 = -0.073534 +0.073557 i	p3 = -222.144 +222.144 i
	p2 = -0.073534 -0.073557 i	p4 = -222.144 -222.144 i
East/West	p1 = -0.073311 +0.073126 i	p3 = -222.144 +222.144 i
	p2 = -0.073311 -0.073126 i	p4 = -222.144 -222.144 i