



REF TEK Bronco

Broadband Seismometer

Date: 2024/4/12

P/N: R-151D-Bronco-120

120 Second

S/N: 247003

Component	Period (sec)	Damping Coefficient	Output Voltage Sensitivity (V/(m/s))	Calibration Sensitivity (m/s ² /A)	Component	Output Voltage Sensitivity (V/(m/s))
U	120.71	0.714	2011.50	10.36	Z	2011.27
W	120.62	0.706	2007.96	10.14	N	2011.16
V	120.42	0.703	2014.36	10.36	E	2011.39

All sensitivities tested at 1 Hz

Zeros: 2 radians

z1 = 0
z2 = 0

Poles: 4 radians

p1 = -0.037024 +0.037024 i p2 = -0.037024 -0.037024 i
p3 = -355.431 +355.431 i p4 = -355.431 -355.431 i

Rev C 2023-09

