



**REF TEK Bronco**

**Broadband Seismometer**

Date: 2024/4/11

P/N: R-151D-Bronco-120

120 Second

S/N: 247008

Component	Period (sec)	Damping Coefficient	Output Voltage Sensitivity (V/(m/s))	Calibration Sensitivity (m/s <sup>2</sup> /A)	Component	Output Voltage Sensitivity (V/(m/s))
U	121.31	0.714	2008.21	10.60	Z	2010.99
W	120.56	0.709	2008.63	10.25	N	2012.39
V	121.21	0.710	2016.14	10.20	E	2009.60

All sensitivities tested at 1 Hz

**Zeros: 2 radians**

z1 = 0  
z2 = 0

**Poles: 4 radians**

p1 = -0.037024 +0.037024 i      p2 = -0.037024 -0.037024 i  
p3 = -355.431 +355.431 i      p4 = -355.431 -355.431 i

Rev C 2023-09

