



**REF TEK Bronco**

**Broadband Seismometer**

Date: 2024/7/22

P/N: R-151D-Bronco-120

120 Second

S/N: 247011

Component	Period (sec)	Damping Coefficient	Output Voltage Sensitivity (V/(m/s))	Calibration Sensitivity (m/s <sup>2</sup> /A)	Component	Output Voltage Sensitivity (V/(m/s))
U	120.99	0.711	2005.20	10.55	Z	2003.46
W	121.14	0.708	2005.21	10.26	N	2002.59
V	121.10	0.712	1999.96	10.35	E	2004.33

All sensitivities tested at 1 Hz

**Zeros: 2 radians**

z1 = 0  
z2 = 0

**Poles: 4 radians**

p1 = -0.037024 +0.037024 i      p2 = -0.037024 -0.037024 i  
p3 = -355.431 +355.431 i      p4 = -355.431 -355.431 i

Rev C 2023-09

