



REF TEK Bronco

Broadband Seismometer

Date: 2024/7/22

P/N: R-151D-Bronco-120

120 Second

S/N: 247016

Component	Period (sec)	Damping Coefficient	Output Voltage Sensitivity (V/(m/s))	Calibration Sensitivity (m/s ² /A)	Component	Output Voltage Sensitivity (V/(m/s))
U	120.94	0.710	2000.67	10.52	Z	1997.92
W	120.72	0.711	1995.45	10.40	N	1996.55
V	120.69	0.708	1997.65	10.32	E	1999.30

All sensitivities tested at 1 Hz

Zeros: 2 radians

z1 = 0
z2 = 0

Poles: 4 radians

p1 = -0.037024 +0.037024 i p2 = -0.037024 -0.037024 i
p3 = -355.431 +355.431 i p4 = -355.431 -355.431 i

Rev C 2023-09

