

DATE: 2025/1/3

REF TEK 147A-01/3 Accelerometer

S/N: 24798

P/N: R-97499-00

NORMALIZED MAGNITUDE - FREQUENCY CHARACTERISTICS

Freq. (Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0004	1.0004	1.0003
10.00	1.0000	1.0000	1.0000
20.00	0.9988	0.9988	0.9991
30.00	0.9969	0.9971	0.9978
40.00	0.9948	0.9950	0.9963
50.00	0.9928	0.9931	0.9951
60.00	0.9913	0.9915	0.9943
70.00	0.9909	0.9909	0.9945
80.00	0.9920	0.9916	0.9961
90.00	0.9950	0.9941	0.9995
100.00	1.0004	0.9988	1.0050
120.00	1.0202	1.0166	1.0239
140.00	1.0534	1.0473	1.0545
160.00	1.0894	1.0805	1.0852

SENSITIVITIES

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.499	2.501
Cross-axis Coupling (g/g)	0.44%	0.32%	0.76%

* g = 9.80 m/s²

Rev H (2020-04)