



REF TEK Bronco

Broadband Seismometer

Date: 2025/2/13

P/N: R-151D-Bronco-120

120 Second

S/N: 247024

Component	Period (sec)	Damping Coefficient	Output Voltage Sensitivity (V/(m/s))	Calibration Sensitivity (m/s ² /A)	Component	Output Voltage Sensitivity (V/(m/s))
U	120.47	0.707	1999.20	10.48	Z	1994.93
W	120.74	0.711	1995.50	10.32	N	1992.80
V	120.77	0.708	1990.10	10.17	E	1997.07

All sensitivities tested at 1 Hz

Zeros: 2 radians

z1 = 0
z2 = 0

Poles: 4 radians

p1 = -0.037024 +0.037024 i p2 = -0.037024 -0.037024 i
p3 = -355.431 +355.431 i p4 = -355.431 -355.431 i

Rev C 2023-09

